







# Product Specification for CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 11

Specification	Sub-Spec	Value
GENERAL FEATURES	Product	Detachable surgical blades
	Blade Type	Detachable Bard Parker
TECHNICAL FEATURES	Disposable blade	Yes
	Size and Shape of blade	As per Cl 5.1 of IS:3319/1995(Latest Revision)
	Commercial Designation of blades as per IS:3319/1995 (Latest Revision)	11
	Tip of blade well defined, central and sharp	Yes
	Cutting edge shall be free from waviness, jags, feathers, nicks and other defects	Yes
	Both the surfaces of blade shall be smooth, and free from tool marks and any sign of corrosion	Yes
	Edges other than the cutting edge shall not be sharp	Yes
	Blade shall fit to the corresponding mount of the handle snugly	Yes
	Blade shall be flat and shall not sprung	Yes
	Thickness of the blade is uniform	Yes
	Both the surfaces of the blade shall be free from roughness and waviness,	Yes
	Hardness of the Stainless blade or Carbon steel blade when tested shall meet the requirement of Cl 7 decimel1 of IS:3319/1995 (Latest Revision)	Yes
	Sterilized blade	Yes
	Method of sterilization	Ionizing (Gamma) radiation process
	Blades shall satisfy the sterility when tested in accordance with IS:10150/1981	Yes
Blades should meet the performance requiremetns when tested as per Cl 9 declmel 2 of IS:3319/1995(Latest Revision)	Yes	
DIMENSIONS, MATERIAL & CONSTRUCTION	Material of Blade	Carbon Steel
	Carbon steel composition for blades	As per Cl 4.1 of IS:3319/1995(Latest Revision)
	Stainless steel designation for blades	Shall conform to X30Cr13 of IS:6911/1992
	Thickness of Blades( mm)	0.37
	Dimensions of blades for small size fittings	As per Table 1 of IS:3319/1995(Latest Revision)
	Dimensions of blades for Large size fittings	As per Table 2 of IS:3319/1995(Latest RA)
PACKING & MARKING	Permanent marking on the Blades	As per Cl 10.1.1 of IS:3319/1995 (Latest Revision)
	Marking on Packages of sterile blades	As per Cl 10.1.3, 10.1.3.1, 10.1.3.2 of IS:3319/1995(Latest Revision)
	Primary packing material for blades	As per Cl 4.3 of IS:3319/1995 (Latest Revision)
	Primary packing & outer Packing of sterile blades	As per Cl 11.2 of IS:3319/1995(Latest Revision)
	Each outer container containing the packages of Sterilized blades shall be provided with sterilization indicator on its outer surface to ensure that the contents of the package have undergone sterilization and to distinguish the sterilization packages from Non-sterile packages	Yes
	Number of blades in a Primary Pack (Nos)	One
	Number of Primary packages in an outer package (Nos)	100
STANDARD	Conformity to standard	IS:3319/1995 (Latest Revision)
	Product is ISI Marked	Yes

	BIS Licence No (CM/L No)	7200055398
CERTIFICATIONS & REPORTS	Availability of Test report from Government / NABL / ILAC accredited Laboratory to prove conforming to the declared specification	Yes
	Test report Number and date	7200055398 31.07.18
	Name and address of Test Lab	BUREAU OF INDIAN STANDARDS, AHMEDABAD
	Certification Available	ISO,EU-CE
	Submission of all the test reports to the buyer along with supplies on demand	Yes
SHELF LIFE / WARRANTY	Warranty (months)	48
	Minimum Shelf life at the time of supply	2/3rd of total shelf life

**Note:** Seller has given an undertaking that it has made arrangements for getting the stores from an authorized distributor / dealer / channel partner of the OEM of the offered product. At the time of delivery of goods, Seller will provide necessary chain documents (in the form of GST Invoice) to prove that the supplied goods are genuine and are being sourced from an authorized distributor / dealer / channel partner of the OEM. In case of any complaint about genuineness of the supplied products, Seller shall be responsible for providing genuine replacement supplies.

**Product Specification for CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 15**

Specification	Sub-Spec	Value
GENERAL FEATURES	Product	Detachable surgical blades
	Blade Type	Detachable Bard Parker
TECHNICAL FEATURES	Disposable blade	Yes
	Size and Shape of blade	As per Cl 5.1 of IS:3319/1995(Latest Revision)
	Commercial Designation of blades as per IS:3319/1995 (Latest Revision)	15
	Tip of blade well defined, central and sharp	Yes
	Cutting edge shall be free from waviness, jags, feathers, nicks and other defects	Yes
	Both the surfaces of blade shall be smooth, and free from tool marks and any sign of corrosion	Yes
	Edges other than the cutting edge shall not be sharp	Yes
	Blade shall fit to the corresponding mount of the handle snugly	Yes
	Blade shall be flat and shall not sprung	Yes
	Thickness of the blade is uniform	Yes
	Both the surfaces of the blade shall be free from roughness and waviness,	Yes
	Hardness of the Stainless blade or Carbon steel blade when tested shall meet the requirement of Cl 7 decimel1 of IS:3319/1995 (Latest Revision)	Yes
	Sterilized blade	Yes
	Method of sterilization	Ionizing (Gamma) radiation process
	Blades shall satisfy the sterility when tested in accordance with IS:10150/1981	Yes
Blades should meet the performance requiremetns when tested as per Cl 9 decimel 2 of IS:3319/1995(Latest Revision)	Yes	
DIMENSIONS, MATERIAL & CONSTRUCTION	Material of Blade	Carbon Steel
	Carbon steel composition for blades	As per Cl 4.1 of IS:3319/1995(Latest Revision)
	Stainless steel designation for blades	Shall conform to X30Cr13 of IS:6911/1992
	Thickness of Blades( mm)	0.37
	Dimensions of blades for small size fittings	As per Table 1 of IS:3319/1995(Latest Revision)
		As per Table 2 of

*M. S. Chillon*  
**M/S. CHILLON'S**  
**Shillong-753018**

*AM*

*Q*

*Inten*

*As per Table 2 of*

	Dimensions of blades for Large size fittings	IS:3319/1995(Latest RA)
PACKING & MARKING	Permanent marking on the Blades	As per Cl 10.1.1 of IS:3319/1995 (Latest Revision)
	Marking on Packages of sterile blades	As per Cl 10.1.3, 10.1.3.1, 10.1.3.2 of IS:3319/1995(Latest Revision)
	Primary packing material for blades	As per Cl 4.3 of IS:3319/1995 (Latest Revision)
	Primary packing & outer Packing of sterile blades	As per Cl 11.2 of IS:3319/1995(Latest Revision)
	Each outer container containing the packages of Sterilized blades shall be provided with sterilization indicator on its outer surface to ensure that the contents of the package have undergone sterilization and to distinguish the sterilization packages from Non-sterile packages	Yes
	Number of blades in a Primary Pack (Nos)	One
	Number of Primary packages in an outer package (Nos)	100
STANDARD	Conformity to standard	IS:3319/1995 (Latest Revision)
	Product is ISI Marked	Yes
	BIS Licence No (CM/L No)	9133569
CERTIFICATIONS & REPORTS	Availability of Test report from Government / NABL / ILAC accredited Laboratory to prove conforming to the declared specification	Yes
	Test report Number and date	128584 DT 12-02-18
	Name and address of Test Lab	SHRIRAMINSTITUTEDELHI
	Certification Available	ISO
	Submission of all the test reports to the buyer along with supplies on demand	Yes
SHELF LIFE / WARRANTY	Warranty (months)	36
	Minimum Shelf life at the time of supply	2/3rd of total shelf life

**Note:** Seller has given an undertaking that it has made arrangements for getting the stores from an authorized distributor / dealer / channel partner of the OEM of the offered product. At the time of delivery of goods, Seller will provide necessary chain documents (in the form of GST Invoice) to prove that the supplied goods are genuine and are being sourced from an authorized distributor / dealer / channel partner of the OEM. In case of any complaint about genuineness of the supplied products, Seller shall be responsible for providing genuine replacement supplies.

## Terms and Conditions

### 1. General Terms and Conditions-

- 1.1 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) as available on the GeM portal (unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable
- 1.2 Terms of delivery: Free Delivery at Site including loading/unloading. In respect of items requiring installation and / or commissioning and other services in the scope of supply (as indicated in respective product category specification / STC / ATC), and the cost of the same is also included in the Contract price.
- 1.2.1 Contracted goods should be delivered at the consignee or designated delivery location as per the working time of the buying organisation. Seller may get the same confirmed from consignee before scheduling delivery.
- 1.2.2 A copy of the contract should be available with the messenger / dispatching agency that delivers the Goods at consignee / delivery location (preferably pasted / attached outside the consignment / package) for easy reference and ease in delivery acceptance.
- 1.3 Delivery period: The Delivery Period/Time shall be essence of the Contract and delivery must be completed not later than such date(s). Any modification thereto shall be mutually agreed and incorporated in the Contract as per the provisions of the GTC.
- 1.4 Performance Security: If the Seller fails or neglects to observe or perform any of his obligations under the contract it shall be liable for the Buyer to forfeit either in whole or in part, the Performance Security furnished by the Seller.

*M*  
 02.04.2018  
 SHRIRAMINSTITUTEDELHI  
 793014

*guthu*

*M*

*S*

*S*

1.5 Taxes and Duties: Contract Prices are all inclusive i.e. including all taxes, duties, local levies / transportation / loading-unloading charges etc. Break up of GST shall be indicated by the Seller while raising invoice / bill on GeM. While submitting the bill / invoice Seller shall undertake that the Goods and Services Tax (GST) charged on this bill is not more than what is payable under the provision on the relevant Act or the Rules made there under and that the Goods on which GST has been charged have not been exempted under the GST Act or the Rules made there under and the charges on account of GST on these goods are correct under the provision of that Act or the rules made there under.

1.6 Octroi Duty and / or other local taxes: Contract Prices are all inclusive hence no reimbursement over and above the contract price(s) shall be allowed to seller towards payment of local taxes (such as levy of town duty, Octroi Duty, Terminal Tax and other levies of local bodies etc).

1.7 Limitation of Liability: The provisions of limitation of liability between Buyer and Seller as given in the GTC shall be applicable here.

1.8 Resolution of disputes: The provisions of DISPUTE RESOLUTION BETWEEN BUYER AND SELLER as given in the GTC shall be applicable here.

1.9 Liquidated Damages: If the Seller fails to deliver any or all of the Goods/Services within the original/re-fixed delivery period(s) specified in the contract, the Buyer will be entitled to deduct/recover the Liquidated Damages for the delay, unless covered under Force Majeure conditions aforesaid, @ 0.5% per week or part of the week of delayed period as pre-estimated damages not exceeding 10% of the contract value without any controversy/dispute of any sort whatsoever. In case, Service Level Agreement (SLA) is applicable the same shall be applicable for the Contract.

1.10 Financial Certificate:


1.10.1 The expenditure involved for this purpose has received the Sanction of the competent financial authority.

1.10.2 The funds are available under the proper head in the sanction budget allotment for the concern financial year.

1.10.3 I have been fully authorized by the department to sign the supply order or incur the liability of the Goods being ordered.

1.11 The bidder should submit a self declaration to the effect in bidder's official letter head that their agency have not been black listed by any Agency whatsoever till date.

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.



Accepted  
REGISTRAR, Shillong-793014



OT/G - 785/2021-22, dt: 18/01/2022

Sanction of the competent authority is hereby conveyed for incurring an expenditure of amount as under towards the cost of Purchase order/Contract placed on the Seller for Supply of Goods/Services as per the contract for making payment to the Seller subject to deduction of TDS as applicable:

Organisation Details	Buyer Details
Type: Central Autonomous Ministry: Ministry of Health and Family Welfare Department: Department of Health and Family Welfare Organisation Name: North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences (NEIGRIHMS) Office Name: Neigrihms, Shillong	<b>A - 03/2019- 20/pt (v)</b> Name: Derenamai Jinthuilu Designation: Jinthuilu Email ID: buycon3.neigrihms.ml@gembuyer.in GSTIN: Address: P.O. NEIGRIHMS, Mawdiangdiang, Shillong EAST KHASI HILLS MEGHALAYA - 793018

**Financial Approval Detail**

Designation of official providing Administration approval:	Director NEIGRIHMS	C-457 (C)/2021-22
IFD Concurrence / Competent Authority (HOD / Head of Office) Approval Required?	YES	C-457 (D)/2021-22
Budget availability	YES	
Designation of official providing Financial approval:	Deputy Director NEIGRIHMS	dt: 4/1/2022

**Seller Details**

Company Name: Yatrah Surgical  
 Email ID: yatrahsurgical222@gmail.com  
 Address: Yatrah Surgical  
 East Delhi  
 DELHI - 110051

**Product Details**

#	Item Description	Model	Ordered Quantity	Unit	Price per Unit inclusive of all Duties and Taxes (in INR)	Total Price (inclusive of all Duties and Taxes (in INR))
1	CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 11	Choice of surgeon, Blades	5000	pieces	1.92	9600.0
2	CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 15	Choise Of Surgeon , Blades	5000	pieces	1.92	9600.0
Total Order Value (in INR)						19200.0

**Consignee Details**

S.No	Consignee	Item	Lot No.	Quantity	Delivery Start After	Delivery To Be Completed By
1	Manly Shabong con22.neigrihms.ml@gembuyer.in P.O. NEIGRIHMS, Mawdiangdiang, Shillong EAST KHASI HILLS MEGHALAYA - 793018	CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 11	-	5000	09-Dec-2021	24-Dec-2021
		CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 15	-	5000	09-Dec-2021	24-Dec-2021

Terms & Conditions

*[Signature]*  
Storekeeper

*[Signature]*  
NEIGRIHMS, Shillong-793018

*[Signature]*

OT/G-786/2021-22, dt: 18/01/2022

Contract



Contract No: GEMC-511687718687595  
Generated Date: 09-Dec-2021

NEIGR/580/A-03/19-20/P4 5

<b>Organisation Details</b> Type: Central Autonomous Ministry: Ministry of Health and Family Welfare Department: Department of Health and Family Welfare Organisation: North Eastern Indira Gandhi Regional Institute of Health and Name: Medical Sciences (NEIGRIHMS) Office Zone: Neigrihms, Shillong	<b>Buyer Details</b> Designation: Jinthulliu Contact No.: -8974990292- Email ID: buycon3.neigrihms.ml@gembuyer.in GSTIN: - Address: P.O. NEIGRIHMS, Mawdiangdiang, Shillong, EAST KHASI HILLS, MEGHALAYA-793018, India
---	---

<b>Financial Approval Detail</b> IFD Concurrence: Yes Designation of Administrative Approval: Director NEIGRIHMS Designation of Financial Approval: Deputy Director NEIGRIHMS	<b>Paying Authority Details</b> Payment Mode: Offline Designation: A Accounts officer Email ID: accounts.neigrihms@nic.in GSTIN: N Address: P.O. NEIGRIHMS, Mawdiangdiang, Shillong, KHASI HILLS EAST, MEGHALAYA-793018, India
--	---

<b>Seller Details</b> GeM Seller ID: E3E3190000929797 Company Name: Yatrah Surgical Contact No.: 09911457723 Email ID: yatrahsurgical222@gmail.com Address: Shop No. 2,A-5,Krishna Nagar,KRISHNA NAGAR, East Delhi, DELHI-110051, - MSME verified: Yes MSME Registration number: DL02D0016505 GSTIN: 07ACQPH7042F1ZK , 07ACQPH7042F1ZK	C-457(E)/2021-22 C-457(G)/2021-22 dt: 4/1/2022
--	--

\*GST / Tax invoice to be raised in the name of - Buyer *- Ra Supply / Billing*



#	Item Description	Category Name & Quadrant	Model	HSN Code	Ordered Quantity	Unit	Lead Time(Days)	Price (Inclusive of all Duties and Taxes in INR)
1	Product Name : CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 22 Brand : CHOICE OF SURGEON Brand Type : Registered Brand Catalogue Status : Catalogue not verified by OEM Selling As : Reseller not verified by OEM	Blades surgical detachable (Q3)	Choice Of Surgeon, Blades	9018	5,000	pieces	-	9,600
2	Product Name : CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 24 Brand : CHOICE OF SURGEON Brand Type : Registered Brand Catalogue Status : Catalogue not verified by OEM Selling As : Reseller not verified by OEM	Blades surgical detachable (Q3)	Choice Of Surgeon, Blades	9018	5,000	pieces	-	9,600
Total Order Value (in INR)								19,200


S.No	Consignee	Item	Lot No.	Quantity	Delivery Start After	Delivery To Be Completed By
1	Designation: Manly Shabong Email ID: con22.neigrihms.ml@gembuyer.in Contact: 0364-2538044- GSTIN: - Address: P.O. NEIGRIHMS, Mawdiangdiang, Shillong, EAST KHASI HILLS, MEGHALAYA 793018, India	CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 22	-	5,000	09-Dec-2021	24-Dec-2021
		CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 24	-	5,000	09-Dec-2021	24-Dec-2021


*Handwritten signatures and notes:*  
 @ Ne Neefur Shis  
 @ DDA / DCA / AO / NW K...  
 NEIGRIHMS Shillong-793018  
 26/01/2022

**Product Specification for CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 22**

Specification	Sub-Spec	Value
GENERAL FEATURES	Product	Detachable surgical blades
	Blade Type	Detachable Bard Parker
TECHNICAL FEATURES	Disposable blade	Yes
	Size and Shape of blade	As per Cl 5.1 of IS:3319/1995(Latest Revision)
	Commercial Designation of blades as per IS:3319/1995 (Latest Revision)	22
	Tip of blade well defined, central and sharp	Yes
	Cutting edge shall be free from waviness, jags, feathers, nicks and other defects	Yes
	Both the surfaces of blade shall be smooth, and free from tool marks and any sign of corrosion	Yes
	Edges other than the cutting edge shall not be sharp	Yes
	Blade shall fit to the corresponding mount of the handle snugly	Yes
	Blade shall be flat and shall not sprung	Yes
	Thickness of the blade is uniform	Yes
	Both the surfaces of the blade shall be free from roughness and waviness.	Yes
	Hardness of the Stainless blade or Carbon steel blade when tested shall meet the requirement of Cl 7 decimal 1 of IS:3319/1995 (Latest Revision)	Yes
	Sterilized blade	Yes
	Method of sterilization	Ionizing (Gamma) radiation process
	Blades shall satisfy the sterility when tested in accordance with IS:10150/1981	Yes
Blades should meet the performance requirements when tested as per Cl 9 decimal 2 of IS:3319/1995(Latest Revision)	Yes	
DIMENSIONS, MATERIAL & CONSTRUCTION	Material of Blade	Carbon Steel
	Carbon steel composition for blades	As per Cl 4.1 of IS:3319/1995(Latest Revision)
	Stainless steel designation for blades	Shall conform to X30Cr13 of IS:6911/1992
	Thickness of Blades( mm)	0.37
	Dimensions of blades for small size fittings	As per Table 1 of IS:3319/1995(Latest Revision)
	Dimensions of blades for Large size fittings	As per Table 2 of IS:3319/1995(Latest RA)
PACKING & MARKING	Permanent marking on the Blades	As per Cl 10.1.1 of IS:3319/1995 (Latest Revision)
	Marking on Packages of sterile blades	As per Cl 10.1.3, 10.1.3.1, 10.1.3.2 of IS:3319/1995(Latest Revision)
	Primary packing material for blades	As per Cl 4.3 of IS:3319/1995 (Latest Revision)
	Primary packing & outer Packing of sterile blades	As per Cl 11.2 of IS:3319/1995(Latest Revision)
	Each outer container containing the packages of Sterilized blades shall be provided with sterilization indicator on its outer surface to ensure that the contents of the package have undergone sterilization and to distinguish the sterilization packages from Non-sterile packages	Yes
	Number of blades in a Primary Pack (Nos)	One
	Number of Primary packages in an outer package (Nos)	100
STANDARD	Conformity to standard	IS:3319/1995 (Latest Revision)
	Product is ISI Marked	Yes

  
  
**SHRI RAMS, Shillong-793016**

  
 [Signature]

  
 [Signature]



	BIS Licence No (CM/L No)	7200055398
CERTIFICATIONS & REPORTS	Availability of Test report from Government / NABL / ILAC accredited Laboratory to prove conforming to the declared specification	Yes
	Test report Number and date	7200055398 31.07.18
	Name and address of Test Lab	BUREAU OF INDIAN STANDARDS, AHMEDABAD
	Certification Available	EU-CE,ISO
	Submission of all the test reports to the buyer along with supplies on demand	Yes
SHELF LIFE / WARRANTY	Warranty (months)	36
	Minimum Shelf life at the time of supply	2/3rd of total shelf life

**Note:** Seller has given an undertaking that it has made arrangements for getting the stores from an authorized distributor / dealer / channel partner of the OEM of the offered product. At the time of delivery of goods, Seller will provide necessary chain documents (in the form of GST Invoice) to prove that the supplied goods are genuine and are being sourced from an authorized distributor / dealer / channel partner of the OEM. In case of any complaint about genuineness of the supplied products, Seller shall be responsible for providing genuine replacement supplies.

**Product Specification for CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 24**

Specification	Sub-Spec	Value
GENERAL FEATURES	Product	Detachable surgical blades
	Blade Type	Detachable Bard Parker
TECHNICAL FEATURES	Disposable blade	Yes
	Size and Shape of blade	As per Cl 5.1 of IS:3319/1995(Latest Revision)
	Commercial Designation of blades as per IS:3319/1995 (Latest Revision)	24
	Tip of blade well defined, central and sharp	Yes
	Cutting edge shall be free from waviness, jags, feathers, nicks and other defects	Yes
	Both the surfaces of blade shall be smooth, and free from tool marks and any sign of corrosion	Yes
	Edges other than the cutting edge shall not be sharp	Yes
	Blade shall fit to the corresponding mount of the handle snugly	Yes
	Blade shall be flat and shall not sprung	Yes
	Thickness of the blade is uniform	Yes
	Both the surfaces of the blade shall be free from roughness and waviness,	Yes
	Hardness of the Stainless blade or Carbon steel blade when tested shall meet the requirement of Cl 7 decimal 1 of IS:3319/1995 (Latest Revision)	Yes
	Sterilized blade	Yes
	Method of sterilization	Ionizing (Gamma) radiation process
	Blades shall satisfy the sterility when tested in accordance with IS:10150/1981	Yes
Blades should meet the performance requirements when tested as per Cl 9 decimal 2 of IS:3319/1995(Latest Revision)	Yes	
DIMENSIONS, MATERIAL & CONSTRUCTION	Material of Blade	Carbon Steel
	Carbon steel composition for blades	As per Cl 4.1 of IS:3319/1995(Latest Revision)
	Stainless steel designation for blades	Shall conform to X30Cr13 of IS:6911/1992
	Thickness of Blades( mm)	0.37
	Dimensions of blades for small size fittings	As per Table 1 of IS:3319/1995(Latest Revision)
		As per Table 2 of

**Shillong**  
**19-10-2018**

*[Handwritten signatures and initials]*

	Dimensions of blades for Large size fittings	IS:3319/1995(Latest RA)
PACKING & MARKING	Permanent marking on the Blades	As per CI 10.1.1 of IS:3319/1995 (Latest Revision)
	Marking on Packages of sterile blades	As per CI 10.1.3, 10.1.3.1, 10.1.3.2 of IS:3319/1995(Latest Revision)
	Primary packing material for blades	As per CI 4.3 of IS:3319/1995 (Latest Revision)
	Primary packing & outer Packing of sterile blades	As per CI 11.2 of IS:3319/1995(Latest Revision)
	Each outer container containing the packages of Sterilized blades shall be provided with sterilization indicator on its outer surface to ensure that the contents of the package have undergone sterilization and to distinguish the sterilization packages from Non-sterile packages	Yes
	Number of blades in a Primary Pack (Nos)	One
	Number of Primary packages in an outer package (Nos)	100
STANDARD	Conformity to standard	IS:3319/1995 (Latest Revision)
	Product is ISI Marked	Yes
	BIS Licence No (CM/L No)	7200055398
CERTIFICATIONS & REPORTS	Availability of Test report from Government / NABL / ILAC accredited Laboratory to prove conforming to the declared specification	Yes
	Test report Number and date	7200055398 31.07.18
	Name and address of Test Lab	BUREAU OF INDIAN STANDARDS, AHMEDABAD
	Certification Available	ISO,EU-CE
	Submission of all the test reports to the buyer along with supplies on demand	Yes
SHELF LIFE / WARRANTY	Warranty (months)	36
	Minimum Shelf life at the time of supply	2/3rd of total shelf life

**Note:** Seller has given an undertaking that it has made arrangements for getting the stores from an authorized distributor / dealer / channel partner of the OEM of the offered product. At the time of delivery of goods, Seller will provide necessary chain documents (in the form of GST Invoice) to prove that the supplied goods are genuine and are being sourced from an authorized distributor / dealer / channel partner of the OEM. In case of any complaint about genuineness of the supplied products, Seller shall be responsible for providing genuine replacement supplies.

## Terms and Conditions

*Handwritten signature*  
**SHRIHMS. SHILONG-793014**

*Handwritten signature*

### 1. General Terms and Conditions-

1.1 This Contract between the Seller and the Buyer, is for the supply of the Goods and/ or Services, detailed in the schedule above, in accordance with the General Terms and Conditions (GTC) as available on the GeM portal (unless otherwise superseded by Goods / Services specific Special Terms and Conditions (STC) and/ or BID/Reverse Auction Additional Terms and Conditions (ATC), as applicable

1.2 Terms of delivery: Free Delivery at Site including loading/unloading. In respect of items requiring installation and / or commissioning and other services in the scope of supply (as indicated in respective product category specification / STC / ATC), and the cost of the same is also included in the Contract price.

1.2.1 Contracted goods should be delivered at the consignee or designated delivery location as per the working time of the buying organisation. Seller may get the same confirmed from consignee before scheduling delivery.

1.2.2 A copy of the contract should be available with the messenger / dispatching agency that delivers the Goods at consignee / delivery location (preferably pasted / attached outside the consignment / package) for easy reference and ease in delivery acceptance.

1.3 Delivery period: The Delivery Period/Time shall be essence of the Contract and delivery must be completed not later than such date(s). Any modification thereto shall be mutually agreed and incorporated in the Contract as per the provisions of the GTC.

1.4 Performance Security: If the Seller fails or neglects to observe or perform any of his obligations under the contract it shall be liable for the

*Handwritten signature*

*Handwritten signature*

Buyer to forfeit either in whole or in part, the Performance Security furnished by the Seller.

1.5 Taxes and Duties: Contract Prices are all inclusive i.e. including all taxes, duties, local levies / transportation / loading-unloading charges etc. Break up of GST shall be indicated by the Seller while raising invoice / bill on GeM. While submitting the bill / invoice Seller shall undertake that the Goods and Services Tax (GST) charged on this bill is not more than what is payable under the provision on the relevant Act or the Rules made there under and that the Goods on which GST has been charged have not been exempted under the GST Act or the Rules made there under and the charges on account of GST on these goods are correct under the provision of that Act or the rules made there under.

1.6 Octroi Duty and / or other local taxes: Contract Prices are all inclusive hence no reimbursement over and above the contract price(s) shall be allowed to seller towards payment of local taxes (such as levy of town duty, Octroi Duty, Terminal Tax and other levies of local bodies etc).

1.7 Limitation of Liability: The provisions of limitation of liability between Buyer and Seller as given in the GTC shall be applicable here.

1.8 Resolution of disputes: The provisions of DISPUTE RESOLUTION BETWEEN BUYER AND SELLER as given in the GTC shall be applicable here.

1.9 Liquidated Damages: If the Seller fails to deliver any or all of the Goods/Services within the original/re-fixed delivery period(s) specified in the contract, the Buyer will be entitled to deduct/recover the Liquidated Damages for the delay, unless covered under Force Majeure conditions aforesaid, @ 0.5% per week or part of the week of delayed period as pre-estimated damages not exceeding 10% of the contract value without any controversy/dispute of any sort whatsoever. In case, Service Level Agreement (SLA) is applicable the same shall be applicable for the Contract.

1.10 Financial Certificate:


1.10.1 The expenditure involved for this purpose has received the Sanction of the competent financial authority.

1.10.2 The funds are available under the proper head in the sanction budget allotment for the concern financial year.


1.10.3 I have been fully authorized by the department to sign the supply order or incur the liability of the Goods being ordered.

1.11 The bidder should submit a self declaration to the effect in bidder's official letter head that their agency have not been black listed by any Agency whatsoever till date.

Note: This is system generated file. No signature is required. Print out of this document is not valid for payment/ transaction purpose.



2024-10-27 10:00:00

  
Biorokopos  
BIOFORMS. Shillong-793014



OT/G-786/2021-22, dt: 18/01/2022

Sanction of the competent authority is hereby conveyed for incurring an expenditure of amount as under towards the cost of Purchase order/Contract placed on the Seller for Supply of Goods/Services as per the contract for making payment to the Seller subject to deduction of TDS as applicable:

Organisation Details	Buyer Details
Type: Central Autonomous Ministry: Ministry of Health and Family Welfare Department: Department of Health and Family Welfare Organisation Name: North Eastern Indira Gandhi Regional Institute of Health and Medical Sciences (NEIGRIHMS) Office Name: Neigrihms, Shillong	<b>A-03/2019-20/pt (v)</b> Name: Dereñamai Jinthuilu Designation: Jinthuilu Email ID: buycon3.neigrihms.ml@gembuyer.in GSTIN: Address: P.O. NEIGRIHMS, Mawdiangdiang, Shillong EAST KHASI HILLS MEGHALAYA - 793018

## Financial Approval Detail

Designation of official providing Administration approval:	Director NEIGRIHMS	C-457(E)/2021-22
IFD Concurrence / Competent Authority (HOD / Head of Office) Approval Required?	YES	
Budget availability	YES	C-457(G)/2021-22
Designation of official providing Financial approval:	Deputy Director NEIGRIHMS	dt: 4/1/2022

## Seller Details

Company Name:	Yatrah Surgical
Email ID:	yatrah surgical222@gmail.com
Address:	Yatrah Surgical East Delhi DELHI - 110051

## Product Details

#	Item Description	Model	Ordered Quantity	Unit	Price per Unit inclusive of all Duties and Taxes (in INR)	Total Price (inclusive of all Duties and Taxes (in INR))
1	CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 22	Choice Of Surgeon, Blades	5000	pieces	1.92	9600.0
2	CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 24	Choice Of Surgeon, Blades	5000	pieces	1.92	9600.0
Total Order Value (in INR)						19200.0

## Consignee Details

S.No	Consignee	Item	Lot No.	Quantity	Delivery Start After	Delivery To Be Completed By
1	Manly Shabong con22.neigrihms.ml@gembuyer.in P.O. NEIGRIHMS, Mawdiangdiang, Shillong EAST KHASI HILLS MEGHALAYA - 793018	CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 22	-	5000	09-Dec-2021	24-Dec-2021
		CHOICE OF SURGEON Detachable surgical blades Carbon Steel Size 24	-	5000	09-Dec-2021	24-Dec-2021

Terms &amp; Conditions

NEIGRIHMS, Shillong - 793018

Jinthuilu

